

一次性口罩成品检验报告

Test report of disposable face masks

QR-8. 2. 6A-02-16-2 №:

报告日期: 2020年08月12日

Report Date: 2020.08.12

检验依据: Q/KK·TC8. 2. 6A-02-16-2020 《一次性防护口罩成品检验规程》

Guide (Provisional), Q/KK·TC8.2.6A-02-16-2020 <Inspection Spec. for the finished products of disposable face masks>

规格: 17.5cm×9.5cm 生产数量: 400000片

抽检数量: 50片

生产批号: 202007 订单号: 2020-KTC115

Size: 17.5cm×9.5cm

Quantity: 400000 PCS

Amount Tested: 50 PCS

LOT : 202007 Order No. : 2020-KTC115

检验项目 Test Item	条款 Terms	性能指标 Standard Requirements	检验结果 Test Result	结论 Conclusion	
基础要求 Basic requirements	1.1	口罩应能安全牢固地护住口、鼻。 The mask should be able to protect the mouth and nose safely and firmly.	符合 Pass	合格 Qualified	
	1.2	口罩原材料不应使用再生料, 无异味。 The raw materials of the masks should not use recycled materials and have no peculiar smell.	符合 Pass	合格 Qualified	
	1.3	口罩不应存在可触及的锐利角和锐利边缘, 不对佩戴者构成伤害。 There should be no sharp angles and sharp edges that can be touched by the mask, which should not cause harm to the wearer.	符合 Pass	合格 Qualified	
	1.4	口罩应便于佩戴和摘除, 在佩戴过程中无明显的压迫感或压痛现象, 对头部活动影响较小。 The mask should be easy to put on and take off. There is no obvious feeling of pressure or tenderness during wearing, which has little effect on head movement.	符合 Pass	合格 Qualified	
外观 Appearance	2.1	口罩表面不应有破损、油污斑渍\变形及其他明显的缺陷。 The surface of the mask should not be damaged, stained, deformed and other obvious defects.	符合 Pass	合格 Qualified	
规格尺寸 Size	3.1	尺寸 17.5×9.5cm 注: 尺寸允许误差±0.5cm。 Size: 17.5cm×9.5cm Note: Tolerance ±0.5cm	17.5cm× 9.3cm	合格 Qualified	
鼻夹 Nose clip	4.1	口罩上应配有鼻夹, 鼻夹长度应不小于8cm。 The nose mask should be equipped with a nose clip, the length of the nose clip should not be less than 8cm.	9.8cm	合格 Qualified	
口罩带 Mask strap	5.1	每根口罩带与口罩体连接点处的断裂强力应不小于10N。 The breaking strength at the connection point between each mask strap and mask should not be less than 10N.	13.5N	合格 Qualified	
微生物清洁度 (cfu/g) Microbial cleanliness	6.1		≤30	21	TYPE II
压力差 (pa/cm ²) Differential pressure	7.1		<40	32.2	TYPE II
细菌过滤 (BFE) % Bacterial filtration efficiency (BFE)	9.1		≥98%	99.1	TYPE II
备注 Remarks	11.1	附照片, 附件一。 Attached photo in Annex			
检验结论 Test conclusion :	经检验的产品样本各项指标均符合《紧急防护民用口罩成品检验规程》的要求及符合谱尼 EN14683:2019 检测。 All indexes of the tested product samples meet the requirements in <Inspection Spec. for the finished products of emergency protective civil masks>, meet the EN14683:2019 test by Pony .				

检验人/Test: 李冬燕

审核人/Check:





合格证
Certificate of Conformity

产品名称: 一次性口罩
Product Name: Disposable Face Mask
规格: 17.5cm x 9.5cm-3p
Specification: 17.5cm x 9.5cm-3p
品牌: TOMASCON
Brand: TOMASCON
主要成分: 65%PP无纺布+35%熔喷布
Materials: 65%PP Nonwoven fabric+35%Melt-blown fabric
执行标准: EN 14683:2019
Reference Standard: EN 14683:2019
型号: KT-N01
Model number: KT-N01
生产批号: 202007
MFI 202007
生产日期: 2020-07-15
MFD 2020-07-15
有效期至: 2023-07-15
Exp 2023-07-15
检验人员: 01
Inspector: 01
建议保存条件: 温度: -10-40℃
Storage: 14 to 104°F (-10 to 40° C)
相对湿度: 85%
Relative humidity: 85%
预期用途: 保护人体免受细菌和病毒等有害物质的侵害, 主要用于保护和隔离。
Intended Use: Protect the human body from harmful substances such as germs and viruses. It is mainly for protection and isolation.

浙江可康医疗科技有限公司
Zhejiang Kekang Medical Technology CO.,LTD.
地址: 浙江省乐清市虹桥镇南村 (乐清市科技创新创业服务中心内)
Address: South Village, Hongqiao Town, Yueqing City, Zhejiang, China (Inside Yueqing Science & Technology Incubation Innovation Service Center)
电话 Tel: 877-61365788

检验合格专用章

合格证
Certificate of Conformity

产品名称: 一次性口罩
Product Name: Disposable Face Mask
规格: 17.5cm x 9.5cm-3p
Specification: 17.5cm x 9.5cm-3p
品牌: TOMASCON
Brand: TOMASCON
主要成分: 65%PP无纺布+35%熔喷布
Materials: 65%PP Nonwoven fabric+35%Melt-blown fabric
执行标准: EN 14683:2019
Reference Standard: EN 14683:2019
型号: KT-N01
Model number: KT-N01
生产批号: 202007
MFI 202007
生产日期: 2020-07-15
MFD 2020-07-15
有效期至: 2023-07-15
Exp 2023-07-15
检验人员: 01
Inspector: 01
建议保存条件: 温度: -10-40℃
Storage: 14 to 104°F (-10 to 40° C)
相对湿度: 85%
Relative humidity: 85%
预期用途: 保护人体免受细菌和病毒等有害物质的侵害, 主要用于保护和隔离。
Intended Use: Protect the human body from harmful substances such as germs and viruses. It is mainly for protection and isolation.

浙江可康医疗科技有限公司
Zhejiang Kekang Medical Technology CO.,LTD.
地址: 浙江省乐清市虹桥镇南村 (乐清市科技创新创业服务中心内)
Address: South Village, Hongqiao Town, Yueqing City, Zhejiang, China (Inside Yueqing Science & Technology Incubation Innovation Service Center)
电话 Tel: 877-61365788

检验合格专用章

